



Ref. Certif. No.

JPTUV-086890

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE**

**CERTIFICAT D'ESSAI OC**

Product  
Produit

Real-Time PCR System

Name and address of the applicant  
Nom et adresse du demandeur

Xi'an Tianlong Science and Technology Co., Ltd.  
No. 389, Zhuhong Road  
Xi'an, Shaanxi Province 710018, P. R. China

Name and address of the manufacturer  
Nom et adresse du fabricant

Xi'an Tianlong Science and Technology Co., Ltd.  
No. 389, Zhuhong Road  
Xi'an, Shaanxi Province 710018, P. R. China

Name and address of the factory  
Nom et adresse de l'usine

Xi'an Tianlong Science and Technology Co., Ltd.  
No. 389, Zhuhong Road  
Xi'an, Shaanxi Province 710018, P. R. China

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

AC 100-240V; 50/60Hz; 900VA

Trademark (if any)  
Marque de fabrique (si elle existe)

TIANLONG

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

N/A

Model / Type Ref.  
Ref. de type

Gentier 96E    Gentier 96S    Gentier 96R    Gentier 96C  
For model differences, refer to the test report.

Additional information (if necessary may also be reported on page 2)  
Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2<sup>ème</sup> page)

A sample of the product was tested and found to be in conformity with  
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 61326-2-6:2012  
IEC 61326-1:2012  
See Test Report for National Differences

As shown in the Test Report Ref. No. which forms part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

50132940 001

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



TÜV Rheinland Japan Ltd.  
Global Technology Assessment Center  
4-25-2 Kita-Yamata, Tsuzuki-ku  
Yokohama 224-0021 Japan  
Phone + 81 45 914-3888  
Fax + 81 45 914-3354  
Mail: info@jpn.tuv.com  
Web: www.tuv.com

Date: 02.04.2018

Signature: Dipl.-Ing. (FH) Fan He

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Real-Time PCR System
Name and address of the applicant	Xi'an Tianlong Science and Technology Co., Ltd. No. 389, Zhuhong Road Xi'an, 710018 Shaanxi, China
Name and address of the manufacturer	Xi'an Tianlong Science and Technology Co., Ltd. No. 389, Zhuhong Road Xi'an, 710018 Shaanxi, China
Name and address of the factory	Xi'an Tianlong Science and Technology Co., Ltd. No. 389, Zhuhong Road Xi'an, 710018 Shaanxi, China
Ratings and principal characteristics	100-240VAC, 50/60Hz, 900VA
Trademark (if any)	TIANLONG
Customer's Testing Facility (CTF) Stage used	N/A
Model / Type Ref.	Gentier 96E , Gentier 96S , Gentier 96R , Gentier 96C For model differences, refer to test report.
Additional information (if necessary may also be reported on page 2)	It was also tested according to IEC 61010-2-010:2014.
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-2-101:2015 for national deviations see test report
As shown in the Test Report Ref. No. which forms part of this Certificate	50086381 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



Date: 05.02.2018

Signature:

Dipl.-Ing. Univ. S. O. Steinke