

**C E R T I F I C A T E**  
of Conformity



**Registration No.:** AK 50378525 0001

**Report No.:** 50079896 001

**Holder:** Xi'an Tianlong Science and  
Technology Co., Ltd.  
No. 389, Zhuhong Road  
Xi'an, Shaanxi Province 710018  
P.R. China

**Product:** Medical Device  
Real-Time PCR System

**Identification:** Type Designation: Gentier 48E , Gentier 48S , Gentier 48R  
Serial No. : n.a.  
Trade Mark : Tianlong  
Remark: Refer to test report 50079896 001 for details.

**Tested acc. to:** IEC 61010-1:2010  
EN 61010-1:2010  
IEC 61010-2-101:2015  
IEC 61326-1:2012  
IEC 61326-2-6:2012

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

**Date** 11.05.2017



**Certification Body**

A handwritten signature in blue ink, likely belonging to Dipl.-Ing. (FH) C. Nasca.

Dipl.-Ing. (FH) C. Nasca

**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**