

C E R T I F I C A T E

of Conformity



Registration No.: AK 50311955 0001

Report No.: 17043135 002

Holder:

Xi'an Tianlong Science and
Technology Co., Ltd.
No. 389, Zhuohong Road
Xi'an, Shaanxi Province 710018
P.R. China

Product:

Medical Device
(Nucleic Acid Extractor)

Identification:

Type Designation: Libex
Serial No. : n.a.
Remark: Refer to test report 17043135 002 for detailed

Tested acc. to:

EN 61010-1:2010
IEC 61010-2-081:2015
IEC 61010-2-101:2015
EN 61326-1:2013
EN 61326-2-6:2013

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Date 16.06.2015



Certification Body

Dipl.-Ing. (FH) C. Nasca

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg