BIOBASE

Biobase Scientific (Shandong) Co.,Ltd ADD: OLABO Intelligent Manufacturing Industrial Park, Ancheng Town, Pingyin County, Jinan City, Shandong

DECLARATION OF CONFORMITY

Technical file of the company mentioned below has been inspected and audit has been completed successfully

2014/35/EU Low Voltage Directive and 2014/30/EU Electromagnetic Compatibility Directive have been taken as reference for these processes

Company Name:

Biobase Scientific (Shandong) Co.,Ltd

OLABO Intelligent Manufacturing Industrial Park, Ancheng

Town, Pingyin County, Jinan City, Shandong

Brand:

BIOBASE

Related Directives and Annex: 2014/35/EU Low Voltage Directive (LVD)

2014/30/EU Electromagnetic Compatibility (EMC)

Related Standards:

EN 61010-1:2010/A1:2019 (Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements) EN IEC 61326-1:2021 (Electrical equipment for measurement, control and

laboratory use - EMC requirements - Part 1: General requirements)

Product(s):

Benchtop pH/ORP/Ion Meter

Type(s)/Model(s):

PH-930

Classification:

Laboratory Equipment

Examination Period:

June 28, 2024

Date of Expiry:

June 27, 2029

Review Result:

We, Biobase Scientific (Shandong) Co., Ltd, declare that during the self-testing and performance evaluation, no Non-compliance according to the requirements of the Low Voltage Directive 2014/35/EU and Electromagnetic Compatibility Directive

2014/30/EU was detected.

Year of DOC marking 2024 器有限公司

RA Specialist : NOXANNE

Document No: BKSD-R24062802